Technical Information

No. FO 4918


Edition: 02/02 - subject to change Supersedes: Initial Release 09/01

Status: valid

XBO[®] R 100 W/45C OFR

■ Product description

- Xenon discharge lamp for DC operation
- · Short arc, high pressure
- Focusing reflector
- Dichroic coating for suppressed IR-output
- For light guides with up to 8 mm diameter
- Cable connection with plug-in contact
- Ozone-free
- Hot re-strikeable



■ Technical data

Rated lamp current	Α	7.2
Rated lamp wattage	W	100
Initial voltage range	V	12 - 14
Ignition voltage (cold/hot)	kV	max. 20 / max. 20
Colour temperature	K	approx. 6000
Colour rendering index	R _a	> 90
Ignition voltage (cold/hot) Colour temperature	kV K R _a	max. 20 / max. 20 approx. 6000



Technical Information

No. FO 4918

Edition: 02/02 - subject to change Supersedes: Initial Release 09/01

Status: valid

XBO® R 100 W/45C OFR

Beam-to-axis angle α	deg	max. 22.4
Length I	mm	max. 77
Length p	mm	max. 6
Reflector diameter D ₁	mm	max. Ø 67.0
Diameter D ₂	mm	Ø 64.0 _{-0,3}
Cable length CL	mm	min. 90
Electrical connectors		Cable with AMP-plug No. 350 809-1

■ Performance data at rated current

Working distance WD	mm	45
Initial aperture lumens	lm	• min. 300 at d = ∅ 3 mm
		• min. 400 at d = ∅ 5 mm
Declared service life 1)	h	500

Mounting

Mounting	rim mounting; for details see Technical Information No. FO 4913	
----------	---	--

Operation conditions

Burning position		horizontal ± 20° (p 20)
Temperature at point A	°C	max. 250 allowed; 175 - 200 optimum
Temperature at point B	°C	max. 350 allowed; 200 - 250 optimum
Cooling		Forced cooling required, air flow velocity and distribution to be chosen in such a way as to ensure proper temperatures at points A and B
Current operation range	Α	min. 7.0, max. 7.4
Power operation range	W	min. 85, max. 115
Polarity		proper polarity to be observed

XBO® R 100 W/45C can be operated either on a standard ballast or on an electronic power supply.

Additional documentation

- Guidelines for Mounting and Centering of Reflector Discharge Lamps (1);
 Technical Information No. FO 4913
- Guidelines for Power Supplies and Ignitors for Xenon Short Arc Lamps Photo Optics No. FO GL-4
- Cooling Scheme; Technical Information No. FO 4935

¹⁾ Operation time after which aperture lumens can decrease to 50% of initial value. Switching cycle 1 hour on, 1 hour off

